


Concise Separations

Coregel ION 310

PN: ICE-99-7752

Processing Method: ION310
 Application: ION310
 Sample Name: Test Mixture
 Column S/N: 12911721

Volume: 20.0 μ l
 Length: 150 mm
 ID: 6.5 mm
 Polymer Lot: 123-74



System 1

Solvent: 0.01 N H₂SO₄

Pressure: Pump A Pressure Profile: Max: 995.8 Min: 981.6

Flow Rate: 0.8 ml/min

Temperature: 30C

RT (min)	Name	k'	Asym	N (EUP)	Res (EUP)	Alpha	S/N	Noise (μ V)
2.04	Sucrose	1.04	1.14	2703	---	---	---	---
2.45	Glucose	1.45	1.05	3000	2.44	1.39	---	---
3.57	Glycerol	2.57	0.99	6194	6.18	1.77	---	---
4.14	Acetic Acid	3.14	0.92	7051	3.03	1.22	---	---
5.37	Ethyl Alcohol	4.37	1.01	7651	5.57	1.39	---	---