

CONCISE SEPARATIONS


Coregel 107H FA

P/N: ICE-99-9867

Processing Method: 107H-FA
Application: 107H-FA
Sample Name: Test Mixture
Column S/N: 12805914

Volume: 20.0 ul
Length: 150 mm
ID: 6.5 mm
Polymer Lot: 056-95

Refractive Index



Flow rate: 0.6mL/min
Temperature: 80C

Solvent: 8.0mM H2SO4

Pump A Pressure Profile: Max: 739.8 Min: 725.5

RT (min)	Name	k'	Asym	N (EUP)	Res (EUP)	Alpha	S/N	Noise (uV)
3.66	Glutaric Acid	365.33	1.39*	4168	---	---	---	---
4.30	Formaldehyde	428.67	1.31*	5753	2.79	1.17	---	---
6.22	Methanol	621.00	1.07	8158	7.66	1.45	---	---
7.07	Ethanol	706.33	1.03	8535	2.93	1.14	---	---

